

Beyond the Visible: Industry 3.91 AFGM 2019 - Cebu (In-Furnace Near IR Borescope)



Mark Bennett¹ -
Neil G. Simpson² - Presenter
Fiona Turner¹



1. AMETEK Land Instruments Ltd,
2. Simpson Combustion and Energy Ltd,



QUALITY CUSTOMER SOLUTIONS




Contents

- Background 2017 GPC summary Cross-Fired Furnace Thermal profile and NOx
- Overview of Equipment Used
 - Need for true temperature measurement at all points
- Original background to end fired project
- Phase 1 Initial 3hr survey
- Equipment rental for extended use
- Thermal and NOx optimisation by customer alone
- Lessons learned and potential for
 - automatic reversal control – reduce energy & emissions
 - Alarm on hot spot deviation and correction
 - Alarm on <1388 NaOH condensation





VIDEO of Continuous Temperature Measurement – Thermocouple Verification










End of Firing L-R

01/09/14

10/10/16

15/03/17










End of Firing R-L

01/09/14

10/10/16

15/03/17





Conclusions from GPC 78th paper

- Operations -Temperature thermal profile
 - Potential energy reduction, pull or yield increase on unrestricted furnace
 - Demonstration of recovery post repair
- Emissions - Flame Intensity and NOx correlation
 - Potential to use as part of NOx/NH3 reduction
- Maintenance – Furnace Sealing
 - Energy and emission reduction with asset protection
- Asset Protection
 - Over and under temperature alarms
 - Identify regenerator restrictions before blockage/deterioration
 - NaOH condensation
 - Preventative maintenance on weak/thin refractory





Thermal Imagers That Accurately Measure Temperatures

- There are two common types of thermal imagers:
 1. Those which produce a thermal image of a scene as a picture or video – known as Non-Radiometric Thermal Imagers. These are commonly used to detect the presence of something hot or cold in a scene like a deer at the side of a road, an intruder or the presence of a hot area somewhere on an object.
 2. Those which produce a thermal image of a scene as a picture or video and accurately measure temperatures anywhere within that thermal scene – known as Radiometric Thermal Imagers. These are commonly used for automation, process control or quality verification.
- Many imagers which are described as Radiometric are only calibrated at one central point in the thermal image and make assumptions about temperature measurement accuracy elsewhere in the scene. It might be argued that one accurate point in the scene makes a thermal imager Radiometric, that isn't what most customers expect.
- When using for control accuracy is required so there is a need test Radiometric Thermal Imagers to verify and document that they accurately measure temperatures across the entire scene.





- NIR-B GLASS thermal imagers are calibrated against blackbody standards over the whole image, not just the central area.
- The Certificate of Conformity shows centre accuracy plus the four corners of the scene.
- Calibration in accordance with ISO 9001:2008 and calibrated against national and international standards.
- Calibration to ISO / IEC 17025:2005 standard.



Instrument Type : ARC 8 11 500 LF
Instrument Serial Number :
Country of Origin : England

Factory Calibration Settings

	Calibration setting ("C")*	Reference standard used
Temperature - Low	9.8	P80P 12102 - 1
Temperature - Medium	248.8	C33 S/N: 20001027
Temperature - High	501.1	C33 S/N: 20001027

*Set with nominal zero error at the temperatures listed. To determine absolute accuracy, measurement uncertainties must also be considered.


Instrument Specification

Range	Wavelength	Field of View (Horizontal)	Focus Range*	Ambient Temperature**	Accuracy**	Frame rate
0 to 500°C	8-14µm	11°	0.5m-infinity	-20 to 60°C	±2% / ±2°C	7.5Hz

*Adjustable **5-95% of range **Specified


Instrument Calibration Results

Image Uniformity Error "C



Reference temperature ("C") = 248.8

Land Instruments International Ltd
Stubley Lane, Dronfield, S18 1D, England
Phone: +44 (0) 1246 417691
Fax: +44 (0) 1246 410585
Email: land.enquiry@ametek.com
www.landinst.com | www.ametek.com



NEED ADDITIONAL SERVICES?




- ISO/IEC 17025 Traceable Calibration
- Technical Product Support
- Annual Spectrum Service Contracts

[VISIT WWW.LANDINST.COM/SERVICES](http://WWW.LANDINST.COM/SERVICES)





Certification Service Coverage

National Accreditation			AMETEK Land Instrumentation		Other Manufacturer's Instrumentation
			Infrared Pyrometers, Thermometers, Thermal Imagers and Scanners	Blackbody Sources	
	United Kingdom	UKAS (United Kingdom Accreditation Service)**	✓ -10 °C to 2500 °C 14 °F to 4532 °F	✓ -10 °C to 1600 °C 14 °F to 2912 °F	✓
	United States	ANAB (ANSI-ASQ National Accreditation Board)	✓ 0 °C to 2300 °C 32 °F to 4172 °F	✓ -10 °C to 1600 °C 32 °F to 2912 °F	✓
	India	NABL (National Accreditation Board for Testing and Calibration Laboratories)	✓ 0 °C to 1600 °C 32 °F to 2912 °F	✗	✓

****Note:** With agreements through the European Cooperation for Accreditation (EA), International Laboratory Accreditation Cooperation (ILCA) and International Accreditation Forum (IAF) - UKAS has recognition in over 90 countries worldwide.





Background to Project

- In early 2018 EU customer with a medium through-put and traditional furnace design had a recent NO_x issue and approached Neil Simpson to investigate options to reduce.
- Several staging options were discussed with decision to focus on baseline and optimisation of existing equipment.
- Part of suggested solution was to use the Ametek Land NIR b to establish possible sources.
- On 6th Feb a 3 hour demonstration was performed.
- The following data was obtained for NO_x and thermal conditions.





Using as a CCTV Standard Batch Line/Movement





B & W Mode 1 Auto Range


Good initial
image



LAND
AMETEK

Date : 06/02/2018

Time : 14:29:45




Emiss : 1.00





B & W Mode 1 Man 1400-1600 Range


Lowering the
temperature
range increase
contrast for
batch
monitoring



LAND
AMETEK


Date : 06/02/2018


Time : 14:30:15



Emiss : 1.00









Reversal – End Firing R-L


Normal view
batch line,
foam line and
front surface



LAND
AMETEK

Date : 06/02/2018

Time : 14:40:50





Emiss : 1.00 -|-





B & W Mode 1 Man 1400- 1600 Range


Increase contrast
to highlight
phase transition
lines.





1.5* B&W1
Man 1400-
1600 Range


Zoom in to
see ripple
lines.



LAND
AMETEK

Date : 06/02/2018

Time : 14:40:50



Emiss : 1.00





1.5* B&W1
Man 1450-1550
Range


Increase
contrast on
front melt.



LAND
AMETEK

Date : 06/02/2018

Time : 14:40:50





Emiss : 1.00





With 20 Degree Isotherms


Add isotherms
to enhance
thermal
difference on
melt





End Firing L-R


Difference in
firing end L-R



LAND
AMETEK

Date : 06/02/2018

Time : 15:01:00



Emiss : 1.00

T.Max 1530
T.Min 1510


T.Max 1510
T.Min 1490

T.Max 1550
T.Min 1530



NaOH Condensation Zones


NaOH
condenses at
1388



LAND
AMETEK

Date : 06/02/2018

Time : 15:01:00



Emiss : 1.00

T.Max 1388
T.Min 1000

T.Max 1400
T.Min 1388


T.Max 1450
T.Min 1400





End L-R 25C Isotherms on Target Wall


Temperatures
seem unusual
since almost
similar.



LAND
AMETEK

Date : 06/02/2018

Time : 15:01:00



Emiss : 1.00

T.Max 1325
T.Min 1300


T.Max 1375
T.Min 1350

T.Max 1350
T.Min 1325



End R-L 25C Isotherms Same Scale


End R-L target
wall temps way
higher



LAND
AMETEK

Date : 06/02/2018

Time : 14:41:00



Emiss : 1.00

T.Max 1325
T.Min 1300


T.Max 1375
T.Min 1350

T.Max 1350
T.Min 1325






End R-L 25C Isotherms +50C



LAND
AMETEK

Date : 06/02/2018

Time : 14:41:00



Emiss : 1.00

T.Max 1375
T.Min 1350


T.Max 1425
T.Min 1400

T.Max 1400
T.Min 1375



End R-L 2* 100C Isotherms


Delta 100C
firing R-L



LAND
AMETEK

Date : 06/02/2018

Time : 14:41:00



Emiss : 1.00

T.Max 1425
T.Min 1325

T.Max 1325
T.Min 1225






End L-R Target
Wall Temps
Same!!

Is there a
regenerator
problem on


LHS?



LAND
AMETEK

Date : 06/02/2018

Time : 15:00:50



Emiss : 1.00



T.Max 1400
T.Min 1300

T.Max 1300
T.Min 1200



P3 End R-L 1.5 zoom



Shows
crown repair
and
possible
flame
licking
touch





P5 End R-L 1.5 zoom


Shows crown
repair. Expansion
joint possible
flame licking
touch stones
and leakage





3* Zoom End
L-R


Looks like a
hole at metal
line.



LAND
AMETEK

Date : 06/02/2018

Time : 15:00:50



Emiss : 1.00





NOx Slides



Beyond the Visible: Industry 3.91 AFGM 2019 - Cebu (In-Furnace Near IR Borescope)



Mark Bennett¹ -
Neil G. Simpson² - Presenter
Fiona Turner¹



1. AMETEK Land Instruments Ltd,
2. Simpson Combustion and Energy Ltd,




QUALITY CUSTOMER SOLUTIONS



1.5 Zoom
NOx mode
1400-1650


Flame very
short
Intense core
+1650



LAND
AMETEK

Date : 06/02/2018

Time : 14:03:01




Emiss : 1.00






1.5 Zoom
NOx mode
1400-1650



LAND
AMETEK

Date : 06/02/2018

Time : 13:47:25



Emiss : 1.00

Long loose
flame with
less intense


flame





1.5 Zoom 1400-1550 NO_x + 50C Isotherms


A lot of flame at
1650-1750



LAND
AMETEK

Date : 06/02/2018

Time : 14:03:09



Emiss : 1.00

T.Max 1750
T.Min 1700

T.Max 1700
T.Min 1650


T.Max 1650
T.Min 1600





1.5 Zoom 1400-1550 NOx + 50C Isotherms


All of flame
below 1700C
majority 1600—
1650C – R-L is
lower NOx



LAND
AMETEK

Date : 06/02/2018

Time : 13:47:25



Emiss : 1.00

T.Max 1750
T.Min 1700

T.Max 1700
T.Min 1650

T.Max 1650
T.Min 1600





Observations From 3 Hour Demonstrations

- There appears to be a problem with the LHS regenerator performance.
- Unbalanced batch pattern
- Instability/problems with combustion air flow
- Significantly higher NO_x firing L-R with corresponding short and intense flame.





Suggested Next Steps

- Address combustion air flow and control
 - Resolved following day
- Investigate regenerator condition and consider cleaning and/or repair
- Consider temporary rental of NIR b to optimise flame conditions.





Phase 2

- Combustion air control and lambda control now operational
- NIR B installed on 21st March
- Regenerator cleaning in progress.





Phase 2 Continued


- After 6 days of furnace optimisation by customer and operators the following NO_x measurements were taken.
- NOTE Neil Simpson NOT involved






27th March
NOx now in
Compliance!

NOx and %O2 During the Reversal






27/3 NOx
1400-1650





27/3 NOx
1400-1650





Conclusion

- In February at 3 hour demonstration the in-balances regenerator temperature and in NOx were obvious.
 - DH side NOx was lower
- Air flow control and lambda probes were then serviced and Regenerator cleaning started.
- On 21st March an NIR b rental unit was installed and thermal profile problems visible.
- By 27th March customer and operators had optimised furnace thermal profile and NOx
- Furnace operating within NOx permit (no additional technology)
 - DH side NOx now higher



Industry 3.9

- **The preceding end-fired case study has shown the ability to interpret the NIR B data to improve the control of the furnace.**
 - Correlation between thermal profile and batch pattern
 - Ability to control thermal profile and hence batch pattern
 - Analysis of target wall temperature indicates issues of potential regenerator blockage
 - additionally ability to use for automatic reversal control for energy and emission reduction
 - Repeatable correlation between peak flame temperature and NOx
 - Identify burner creating problem and enable optimisation
- **The need to have accurate, calibrated temperatures for control**



